

## Certificate of analysis

<b>Product:</b>	<i>Agrobacterium Tumefaciens</i>
<b>Catalog No:</b>	Oneq-A58945-100D
<b>Lot No:</b>	Oneq-A58945-100D110326

### Storage Conditions:

All *Agrobacterium Tumefaciens* PCR detection kit reagents are shipped ambiently. Expiration date is 03-19-2028. *Agrobacterium Tumefaciens* PCR detection kit reagents are stable at room temperature.

### Specifications:

Reagents: Universal qPCR Master Mix	Cat No: Oneq-A58945-100D
Qty: 1000 µL	Lot No: Oneq-A58945-100D110326
<b>Passed</b>	

Reagents: Primer, Probes and Internal Control Universal Mix	Cat No: Oneq-A58945-100D
Qty: 72 µL	Lot No: Oneq-A58945-100D110326
<b>Passed</b>	

Reagents: Positive Control	Cat No: Oneq-A58945-100D
Qty: 80 µL	Lot No: Oneq-A58945-100D110326
<b>Passed</b>	

Reagents: Negative Control	Cat No: Oneq-A58945-100D
Qty: 300 µL	Lot No: Oneq-A58945-100D110326
<b>Passed</b>	

Reagents: PCR grade Water Control	Cat No: Oneq-A58945-100D
Qty: 1000 µL	Lot No: Oneq-A58945-100D110326
<b>Passed</b>	

### **Reagent Performance**

All reagents are stable at room temperature. All new lots of reagents are tested with the kit. In order to pass quality control tests the reagents must yield within 10% of the DNA of the previously passed lots.

### **Conclusions:**

This lot of *Agrobacterium Tumefaciens* PCR detection kit are released for product sales.

### **Technical assistance:**

The manuals of the web page are only referential, only use the manuals sent to your email. For consultation on the use, contact us [info@bioingentech.com](mailto:info@bioingentech.com)

### **Certificate validated by:**

**BQ. Benjamín Catalán Soto - BQ. Rodrigo Iturra Barra**  
**Quality Control Managers**