

Certificate of analysis

Product:	<i>Botryosphaeriaceae</i>
Catalog No:	Oneq-A597-100D
Lot No:	Oneq-A597-100D110326

Storage Conditions:

All *Botryosphaeriaceae* PCR detection kit reagents are shipped ambiently.
Expiration date is 03-19-2028. *Botryosphaeriaceae* PCR detection kit reagents are stable at room temperature.

Specifications:

Reagents: Universal qPCR Master Mix	Cat No: Oneq-A597-100D
Qty: 500 µL	Lot No: Oneq-A597-100D110326
	Passed
Reagents: Primer, Probes and Internal Control Universal Mix	Cat No: Oneq-A597-100D
Qty: 36 µL	Lot No: Oneq-A597-100D110326
	Passed
Reagents: Positive Control	Cat No: Oneq-A597-100D
Qty: 40 µL	Lot No: Oneq-A597-100D110326
	Passed
Reagents: Negative Control	Cat No: Oneq-A597-100D
Qty: 150 µL	Lot No: Oneq-A597-100D110326
	Passed
Reagents: PCR grade Water Control	Cat No: Oneq-A597-100D
Qty: 500 µL	Lot No: Oneq-A597-100D110326
	Passed

Reagent Performance

All reagents are stable at room temperature. All new lots of reagents are tested with the kit. In order to pass quality control tests the reagents must yield within 10% of the DNA of the previously passed lots.

Conclusions:

This lot of *Botryosphaeriaceae* PCR detection kit are released for product sales.

Technical assistance:

The manuals of the web page are only referential, only use the manuals sent to your email. For consultation on the use, contact us info@bioingentech.com

Certificate validated by:

BQ. Benjamín Catalán Soto - BQ. Rodrigo Iturra Barra

Quality Control Managers