

Certificate of analysis

Product:	<i>Candidatus Liberibacter Asiaticus</i>
Catalog No:	Oneq-A598-100D
Lot No:	Oneq-A598-100D110326

Storage Conditions:

All *Candidatus Liberibacter Asiaticus* PCR detection kit reagents are shipped ambiently. Expiration date is 03-19-2028. *Candidatus Liberibacter Asiaticus* PCR detection kit reagents are stable at room temperature.

Specifications:

Reagents: Universal qPCR Master Mix	Cat No: Oneq-A598-100D
Qty: 1000 µL	Lot No: Oneq-A598-100D110326
Passed	

Reagents: Primer, Probes and Internal Control Universal Mix	Cat No: Oneq-A598-100D
Qty: 72 µL	Lot No: Oneq-A598-100D110326
Passed	

Reagents: Positive Control	Cat No: Oneq-A598-100D
Qty: 80 µL	Lot No: Oneq-A598-100D110326
Passed	

Reagents: Negative Control	Cat No: Oneq-A598-100D
Qty: 300 µL	Lot No: Oneq-A598-100D110326
Passed	

Reagents: PCR grade Water Control	Cat No: Oneq-A598-100D
Qty: 1000 µL	Lot No: Oneq-A598-100D110326
Passed	

Reagent Performance

All reagents are stable at room temperature. All new lots of reagents are tested with the kit. In order to pass quality control tests the reagents must yield within 10% of the DNA of the previously passed lots.

Conclusions:

This lot of *Candidatus Liberibacter Asiaticus* PCR detection kit are released for product sales.

Technical assistance:

The manuals of the web page are only referential, only use the manuals sent to your email. For consultation on the use, contact us info@bioingentech.com

Certificate validated by:

BQ. Benjamín Catalán Soto - BQ. Rodrigo Iturra Barra
Quality Control Managers