**BioinGentech**
**O'Higgins 1213, Concepción, Chile**
**www.bioingentech.com**
**info@bioingentech.com**

# Certificate of analysis

| | |
|---|---|
| **Date:** | 07/31/2026 |
| **Product:** | ***Infectious Bursal Disease A3*** |
| **Catalog No:** | Oneq-V892-150D |
| **Lot No:** | Oneq-V892-150D072226 |

**Storage Conditions:**

All **Infectious Bursal Disease A3.** PCR detection kit reagents are shipped ambiently. Expiration date is 07-31-2028. ***Infectious Bursal Disease A3*** PCR detection kit reagents are stable at room temperature.

**Specifications:**

| | |
|---|---|
| Reagents: Universal qPCR Master Mix | Cat No: Oneq-V892-150D |
| Qty: 1500 µL | Lot No: Oneq-V892-150D072226 |
| | **Passed** |

| | |
|---|---|
| Reagents: Primer, Probes and Internal Control Universal Mix | Cat No: Oneq-V892-150D |
| Qty: 108 µL | Lot No: Oneq-V892-150D072226 |
| | **Passed** |

| | |
|---|---|
| Reagents: Positive Control | Cat No: Oneq-V892-150D |
| Qty: 120 µL | Lot No: Oneq-V892-150D072226 |
| | **Passed** |

| | |
|---|---|
| Reagents: Negative Control | Cat No: Oneq-V892-150D |
| Qty: 450 µL | Lot No: Oneq-V892-150D072226 |
| | **Passed** |

| | |
|---|---|
| Reagents: PCR grade Water Control | Cat No: Oneq-V892-150D |
| Qty: 1500 µL | Lot No: Oneq-V892-150D072226 |
| | **Passed** |

| | |
|---|---|
| Reagents: Exogenous RNA Control | Cat No: Oneq-V892-150D |
| Qty: 105 µL | Lot No:Oneq-V892-150D072226 |
| | **Passed** |

| | |
|---|---|
| Reagents: Exogenous RNA control Primer and Probes Mix | Cat No: Oneq-V892-150D |

**BioinGentech**
**O'Higgins 1213, Concepción, Chile**
**www.bioingentech.com**
**info@bioingentech.com**

| | |
|---|---|
| Qty: 112.5 µL | Lot No:Oneq-V892-150D072226 |
| | **Passed** |

| | |
|---|---|
| Reagents: One qRTPCR Enzyme Mix | Cat No:Oneq-V892-150D |
| Qty: 110 µL | Lot No:Oneq-V892-150D072226 |
| | **Passed** |

**Reagent Performance**

All reagents are stable at room temperature. All new lots of reagents are tested with the kit. In order to pass quality control tests the reagents must yield within 10% of the DNA of the previously passed lots.

**Conclusions:**

This lot of ***Infectious Bursal Disease A3*** PCR detection kit are released for product sales.

**Technical assistance:**

The manuals of the web page are only referential, only use the manuals sent to your email. For consultation on the use, contact us info@bioingentech.com

**Certificate validated by:**

**BQ. Benjamín Catalán Soto - BQ. Rodrigo Iturra Barra**

**Quality Control Managers**