

## Certificate of analysis

<b>Product:</b>	lactobacillus spp
<b>Catalog No:</b>	Oneq-F566-100D
<b>Lot No:</b>	Oneq-F566-100D23082025

### Storage Conditions:

All lactobacillus spp PCR detection kit reagents are shipped ambiently.

Expiration date is 23-08-2026. lactobacillus spp KIT PCR detection kit reagents are stable at room temperature.

### Specifications

Reagents: Universal qPCR Master Mix	Cat No: Oneq-F566-100D
Qty: 1000 µL	Lot No: Oneq-F566-100D23082025
<b>Passed</b>	
Reagents: Primer, Probes and Internal Control Universal Mix	Cat No: Oneq-F566-100D
Qty: 72 µL	Lot No: Oneq-F566-100D23082025
<b>Passed</b>	
Reagents: Positive Control	Cat No: Oneq-F566-100D
Qty: 80 µL	Lot No: Oneq-F566-100D23082025
<b>Passed</b>	
Reagents: Negative Control	Cat No: Oneq-F566-100D
Qty: 300 µL	Lot No: Oneq-F566-100D23082025
<b>Passed</b>	
Reagents: PCR grade Water Control	Cat No: Oneq-F566-100D
Qty: 1000 µL	Lot No: Oneq-F566-100D23082025
<b>Passed</b>	

### Reagent Performance

Versión 02

23/08/2025

All reagents are stable at room temperature. All new lots of reagents are tested with the kit. In order to pass quality control tests the reagents must yield within 10% of the DNA of the previously passed lots.

**Conclusions:**

This lot of *Lactobacillus* spp PCR detection kit are released for product sales.

**Technical assistance:**

The manuals of the web page are only referential, only use the manuals sent to your email. For consultation on the use, contact us [info@kitpcr.com](mailto:info@kitpcr.com)

**Certificate validated by:**

**BQ. Diego Alonso Abarzua Montanares**

**BQ. Benjamín Ignacio Catalán Soto**