

## Certificate of analysis

|                    |                                        |
|--------------------|----------------------------------------|
| <b>Product:</b>    | Middle East Respiratory Syndrome (UpE) |
| <b>Catalog No:</b> | Oneq-VH597-100R                        |
| <b>Lot No:</b>     | Oneq-VH597-100R15072025                |

### Storage Conditions:

All Middle East Respiratory Syndrome (UpE) PCR detection kit reagents are shipped ambiently.

Expiration date is 15-07-2026, Middle East Respiratory Syndrome (UpE) PCR detection kit reagents are stable at room temperature.

### Specifications:

|                                                             |                               |
|-------------------------------------------------------------|-------------------------------|
| Reagents: Universal qPCR Master Mix                         | Cat No: Oneq-VH597-100R       |
| Qty: 1000 µL                                                | LotNo:Oneq-VH597-100R15072025 |
|                                                             | <b>Passed</b>                 |
| Reagents: Primer, Probes and Internal Control Universal Mix | Cat No: Oneq-VH597-100R       |
| Qty: 72 µL                                                  | LotNo:Oneq-VH597-100R15072025 |
|                                                             | <b>Passed</b>                 |
| Reagents: Positive Control                                  | Cat No: Oneq-VH597-100R       |
| Qty: 80 µL                                                  | LotNo:Oneq-VH597-100R15072025 |
|                                                             | <b>Passed</b>                 |
| Reagents: Negative Control                                  | Cat No: Oneq-VH597-100R       |
| Qty: 300 µL                                                 | LotNo:Oneq-VH597-100R15072025 |
|                                                             | <b>Passed</b>                 |
| Reagents: PCR grade Water Control                           | Cat No: Oneq-VH597-100R       |
| Qty: 1000 µL                                                | LotNo:Oneq-VH597-100R15072025 |
|                                                             | <b>Passed</b>                 |
| Reagents: Exogenous RNA Control                             | Cat No: Oneq-VH597-100R       |

|                                                       |                               |
|-------------------------------------------------------|-------------------------------|
| Qty: 70 µL                                            | LotNo:Oneq-VH597-100R15072025 |
| <b>Passed</b>                                         |                               |
| Reagents: Exogenous RNA control Primer and Probes Mix | Cat No: Oneq-VH597-100R       |
| Qty: 75 µL                                            | LotNo:Oneq-VH597-100R15072025 |
| <b>Passed</b>                                         |                               |
| Reagents: One qRTPCR Enzyme Mix                       | Cat No: Oneq-VH597-100R       |
| Qty: 70 µL                                            | LotNo:Oneq-VH597-100R15072025 |
| <b>Passed</b>                                         |                               |

### Reagent Performance

All reagents are stable at room temperature. All new lots of reagents are tested with the kit. In order to pass quality control tests the reagents must yield within 10% of the DNA of the previously passed lots.

### Conclusions:

This lot of Middle East Respiratory Syndrome (UpE) PCR detection kit are released for product sales.

### Technical assistance:

The manuals of the web page are only referential, only use the manuals sent to your email. For consultation on the use, contact us [info@kitpcr.com](mailto:info@kitpcr.com)

Certificate validated by:

BQ. Diego Alonso Abarzua Montanares

BQ. Benjamín Ignacio Catalán Soto