

Certificate of analysis

Product:	roseburia spp
Catalog No:	Oneq-H746-100D
Lot No:	Oneq-H746-100D26082025

Storage Conditions:

All roseburia spp PCR detection kit reagents are shipped ambiently.

Expiration date is 26-08-2026. roseburia spp PCR detection kit reagents are stable at room temperature.

Specifications

Reagents: Universal qPCR Master Mix	Cat No: Oneq-H746-100D
Qty: 1000 µL	LotNo: Oneq-H746-100D26082025
Passed	
Reagents: Primer, Probes and Internal Control Universal Mix	Cat No: Oneq-H746-100D
Qty: 72 µL	LotNo: Oneq-H746-100D26082025
Passed	
Reagents: Positive Control	Cat No: Oneq-H746-100D
Qty: 80 µL	LotNo: Oneq-H746-100D26082025
Passed	
Reagents: Negative Control	Cat No: Oneq-H746-100D
Qty: 300 µL	LotNo: Oneq-H746-100D26082025
Passed	
Reagents: PCR grade Water Control	Cat No: Oneq-H746-100D
Qty: 1000 µL	LotNo: Oneq-H746-100D26082025
Passed	

Reagent Performance

Versión 02

26/08/2025

All reagents are stable at room temperature. All new lots of reagents are tested with the kit. In order to pass quality control tests the reagents must yield within 10% of the DNA of the previously passed lots.

Conclusions:

This lot of roseburia spp PCR detection kit are released for product sales.

Technical assistance:

The manuals of the web page are only referential, only use the manuals sent to your email. For consultation on the use, contact us info@kitpcr.com

Certificate validated by:

BQ. Diego Alonso Abarzua Montanares

BQ. Benjamín Ignacio Catalán Soto