



Quality Control




Dilution	1	2	3	4	5	6	7	8
Concentration	1×10^{-1}	1×10^{-2}	1×10^{-3}	1×10^{-4}	1×10^{-5}	1×10^{-6}	1×10^{-7}	1×10^{-8}
N° of copies	1×10^8	1×10^7	1×10^6	1×10^5	1×10^4	1×10^3	1×10^2	1×10^2



Target Name	Sample Name	Reporter	Ct values
<i>Babesia bovis</i>	1	FAM	14,43
	2	FAM	18,51
	3	FAM	21,76
	4	FAM	24,12
	5	FAM	27,87
	6	FAM	30,54
	7	FAM	33,97
	8	FAM	35,42
	Negative Control	FAM	Undetermined





Certificate validated by:

BQ. Benjamín Catalán Soto - BQ. Rodrigo Iturra Barra
Quality Control Managers