



Quality Control




Dilution	1	2	3	4	5	6	7
Concentration	1×10^{-1}	1×10^{-2}	1×10^{-3}	1×10^{-4}	1×10^{-5}	1×10^{-6}	1×10^{-7}
Nº of copies	1×10^7	1×10^6	1×10^5	1×10^4	1×10^3	1×10^2	1×10^1



Target Name	Sample Name	Reporter	Ct values
<i>Bifidobacterium</i>	1	ROX	12,43
	2	ROX	15,67
	3	ROX	20,92
	4	ROX	23,72
	5	ROX	26,61
	6	ROX	29,84
	7	ROX	32,63
	Negative Control	ROX	Undetermined





Certificate validated by:

BQ. Benjamín Catalán Soto - BQ. Rodrigo Iturra Barra

Quality Control Managers