



Quality Control




Dilution	1	2	3	4	5	6	7
Concentration	1×10^{-1}	1×10^{-2}	1×10^{-3}	1×10^{-4}	1×10^{-5}	1×10^{-6}	1×10^{-7}
N° of copies	1×10^7	1×10^6	1×10^5	1×10^4	1×10^3	1×10^2	1×10^1



Target Name	Sample Name	Reporter	Ct values
<i>Bovine Viral Diarrhea Virus</i>	1	HEX	15,76
	2	HEX	16,29
	3	HEX	19,42
	4	HEX	22,56
	5	HEX	25,89
	6	HEX	27,85
	7	HEX	32,86
	Negative Control	HEX	Undetermined





Certificate validated by:

BQ. Benjamín Catalán Soto - BQ. Rodrigo Iturra Barra

Quality Control Managers