



Quality Control



Dilution	1	2	3	4	5	6	7	8
Concentration	1×10^{-1}	1×10^{-2}	1×10^{-3}	1×10^{-4}	1×10^{-5}	1×10^{-6}	1×10^{-7}	1×10^{-8}
N° of copies	1×10^8	1×10^7	1×10^6	1×10^5	1×10^4	1×10^3	1×10^2	1×10^1



Target Name	Sample Name	Reporter	Ct values
<i>Mycoplasma synoviae</i>	1	FAM	11,33
	2	FAM	15,57
	3	FAM	18,67
	4	FAM	21,62
	5	FAM	25,18
	6	FAM	28,49
	7	FAM	31,32
	8	FAM	34,21
	Negative Control	FAM	Undetermined





Certificate validated by:

BQ. Benjamín Catalán Soto - BQ. Rodrigo Iturra Barra

Quality Control Managers