



Quality Control




Dilution	1	2	3	4	5	6	7
Concentration	1×10^{-1}	1×10^{-2}	1×10^{-3}	1×10^{-4}	1×10^{-5}	1×10^{-6}	1×10^{-7}
Nº of copies	1×10^8	1×10^7	1×10^6	1×10^5	1×10^4	1×10^3	1×10^2



Target Name	Sample Name	Reporter	Ct values
<i>Newcastle disease virus</i>	1	FAM	15,56
	2	FAM	18,98
	3	FAM	22,36
	4	FAM	25,22
	5	FAM	28,39
	6	FAM	32,17
	7	FAM	35,23
	Negative Control	FAM	Undetermined





Certificate validated by:

BQ. Benjamín Catalán Soto - BQ. Rodrigo Iturra Barra

Quality Control Managers