

PCR Kit Quality Control Certificate

Pathogen
<i>Puccinia striiformis f. sp. Tritici</i>

Table 1: Concentration in ng/μL of the sample control and its dilutions.

	1	2	3	4	5	6	7	8	9
	1×10^{-1}	1×10^{-2}	1×10^{-3}	1×10^{-4}	1×10^{-5}	1×10^{-6}	1×10^{-7}	1×10^{-8}	1×10^{-9}
N° of copies	1×10^8	1×10^7	1×10^6	1×10^5	1×10^4	1×10^3	1×10^2	1×10^1	1×10^0




Figure 1: Logarithmic amplification curves for *Puccinia striiformis f. sp. Tritici*




Figure 2: Standard amplification curves for *Puccinia striiformis f. sp. Tritici*




Figure 3: Standard curve for *Puccinia striiformis f. sp. Tritici*

Summary of Ct values

Table 2: Ct values obtained for each dilution of the sample control.

Target Name	Sample Name	Reporter	Ct values
<i>Puccinia striiformis f. sp. Tritici</i>	1	FAM	14,17
	2	FAM	17,13
	3	FAM	20,31
	4	FAM	23,41
	5	FAM	26,41
	6	FAM	29,93
	7	FAM	32,98
	8	FAM	36,01
	Negative Control	FAM	Undetermined